


**FREQUENCY PROFILE  
USING AN HP 5345A  
ELECTRONIC FREQUENCY COUNTER  
AND AN HP 5359A  
TIME SYNTHESIZER**


HP-IB



## INTRODUCTION


The measurement power of modern day electronic instruments can be further enhanced by operation in the remote mode. Instruments designed to operate on the Hewlett-Packard Interface bus (HP-IB), which is an implementation of IEEE Standard 488-1978, can be very easily interfaced by means of a simple standardized cable interconnection. A computing controller like the Hewlett-Packard 9825A may then be used to control instrument operation. Complex measurements and data collection that would require much interaction by a skilled user under manual operation can now be performed automatically in an HP-IB system.

Frequency Profile is the time variation of frequency within a signal. This Application Note describes the use of an HP 5345A Electronic Counter, an HP 5359A Time Synthesizer together with an HP 9825A Calculator, to obtain the frequency profile of an unknown signal. Two measurement techniques are used, the "time" and "events" modes. In the time mode of operation, the external gate of the counter is controlled by a selectable width and delay signal, generated by the time synthesizer. This gate pulse can move inside a continuous wave stream or burst to obtain its frequency profile. (Figure 1)



Time Mode of Operation  
**Figure 1**

When in the events mode of operation, the width of the gate signal is the period of one event and can be programmed to move inside a continuous wave or burst to capture any one particular event. The period of each event can be different but the system can still capture its characteristics. (Figure 2)



Events Mode of Operation  
**Figure 2**

## CONFIGURATION BLOCK DIAGRAM




Figure 3

## PROGRAM DESCRIPTION

The overall program is divided in two sections, the Time Operation and Events Operation section. Each one is described as follows:

### TIME OPERATION

The equipment set-up is as shown in Figure 3 with the exception of the EVENTS INPUT to the 5359A which is not used in this mode of operation. It is essential that the signal source have a synchronous trigger capable of driving the external trigger input of the 5359A Time Synthesizer. The 5359A can externally trigger on either the positive or negative going slope of the external trigger. Typical timing relationships for this mode of operation are shown in Figure 4.




Figure 4

- Width + Delay  $\leq 160\text{ms}$
- SYNC DELAY in PRESET. Insertion Delay  $\leq 140\text{ nsec}$

The calculator will allow the user to select the 5359A's output width, delay and delay step size. It will then ask if either automatic or manual operation is desired. In manual operation, the delay can be incremented one step at a time, thereby allowing the gate pulse to walk through the input signal stream. After each step, the counter will make a frequency measurement and the calculator will print out the step number and frequency measured. Automatic operation is similar, except that the calculator will automatically step through the input signal stream a certain number of steps as specified by the user.

When in Time Operation, the 5345A is programmed for an internal gate of 1 second and an external gate as set by the 5359A. The effective gate time is the sum of individual external gates adding up to the 1 sec internal gate. This technique is called **FREQUENCY AVERAGING** and its benefit is increased resolution and accuracy.

## EVENTS OPERATION

Figure 3 shows the equipment set up for EVENTS OPERATION. As is the case for TIME OPERATION the signal source must have a synchronous trigger capable of driving the external trigger input of the 5359A Time Synthesizer. Typical timing relationships for this mode of operation are shown in Figure 5.




Figure 5

- Width + Delay  $\leq 16777215$  Events.

The program will first calibrate itself and frame the second event in the input signal stream. The counter will make a frequency and period measurement and the calculator will print out the results. The calculator will then allow a user to select any event/slope of the input signal stream. Slope selection can vary an event count as shown in Figure 6. As can be seen, event number 3 negative slope is further in the pulse stream than event number 3 positive slope.




Figure 6

## SYSTEM PERFORMANCE CHARACTERISTICS

### A. TIME MODE OF OPERATION

For Time Mode of Operation, system resolution and accuracy can be computed as follows:

$$1. \text{ Resolution (Hz)} = \pm \left[ \frac{2 \text{ nsec}}{\text{Ext Gate} \times \sqrt{N}} \times \text{freq} \right] \pm \left[ \frac{\frac{\text{trigger error}}{\text{Ext Gate}}}{\sqrt{N}} \times \text{freq} \right]$$

where:

$$N = \frac{\text{Internal Gate}^*}{\text{External Gate}}$$

\*Program sets Internal Gate to 1 sec. If another internal gate time is desired, change line 35 of the program to reflect this change.

$$\text{trigger error (rms)} = \frac{1.4 \times \sqrt{X^2 + en^2}}{\text{Input Slew Rate at trigger point}}$$

$X = 300 \mu V$  for 5345A. Effective rms noise of counter's input channel.

$en$  = rms noise voltage of input signal measured over a bandwidth equal to the counter's bandwidth.

$$2. \text{ Accuracy (Hz)} = \pm \text{Resolution} \pm [5345A \text{ time base error} \times \text{freq.}] \pm \text{Bias}$$

### 5345A Time Base:

Standard Time Base: High Stability 10544A 10 MHz Oven Oscillator

Aging Rate:  $< 5 \times 10^{-10}$  per day

Short Term:  $< 1 \times 10^{-11}$  for 1 sec average

Temperature:  $< 7 \times 10^{-9}$ ,  $0^\circ C$  to  $55^\circ C$

Line Voltage:  $< 1 \times 10^{-10}$ , +10% from nominal

Option 001 Time Base: Voltage Controlled 10 MHz Room Temperature Crystal

Aging Rate:  $< 3 \times 10^{-7}$  per month

Short Term:  $< 2 \times 10^{-9}$  rms for 1 sec

Temperature:  $< 2 \times 10^{-6}$ ,  $25^\circ C$  to  $35^\circ C$

Line Voltage:  $< 5 \times 10^{-6}$ ,  $0^\circ C$  to  $55^\circ C$

Line Voltage:  $< 1 \times 10^{-8}$ , +10% from nominal

### Bias:

While displaying a relatively large amount of very usable resolution, the absolute accuracy of frequency average measurements will be less than the displayed resolution (will read slightly low) due to known circuit delays within the counter. The gate time error is a fixed time error of a few pico seconds; therefore, the frequency error will become more pronounced as the EXT GATE signal becomes narrower. For any given EXT GATE width a good portion of this error can be calibrated out of the measurement when maximum accuracy is required.

When averaging 100 measurements for an improvement of X10 in stable resolution the gate error is usually insignificant.

When taking more samples for greater resolution, greater absolute accuracy can also be obtained by using a calibration factor for the particular EXT GATE width being used. The calibration factor may be determined by counting a stable high frequency signal applied to CHAN A using the FREQ A mode and the same external gating pattern as used for the signal of interest. Average as many readings as necessary to get the resolution desired. When using the 10 MHz STANDARD available on the rear panel of the 5345A the counter may display 9.99xx MHz instead of the expected 10.0000 MHz. The calibration factor, cf, for this external gate width will be

$$Cf = \frac{\text{known frequency}}{\text{displayed frequency}} = \frac{10.000 \text{ MHz}}{9.99xx \text{ MHz}} = 1.00xx$$

$$\text{The unknown pulse carrier} = \text{displayed pulse carrier} \times Cf.$$

If the 10 MHz STANDARD OUTPUT of the counter is used to determine the calibration factor, the absolute accuracy of a FREQUENCY AVERAGED pulse measurement can be improved by a factor of at least 20 over the single shot case (where accuracy is essentially the same as the resolution).

#### EXAMPLE:

Video signal applied to mainframe = 345.678910 MHz  
External gate pulse width = 1  $\mu$  sec (or 500 clock pulses)  
Single shot video resolution = 345 MHz  
Single shot accuracy =  $\frac{\pm 1 \text{ clock pulse}}{500 \text{ clock pulses}} = 0.2\%$

#### TAKING A MEASUREMENT

RMS resolution when averaging =  $\sqrt{N}$ , so for a visibly stable extra digit take 1000 samples.

Measurement with 1000 samples  $F_{\text{VIDEO}} = 345.5$  Measurement is within  $\pm 1$  count of actual value.

Measurement with 100,000 samples  $F_{\text{VIDEO}} = 345.57$  Resolution is stable but actual value is low.

#### DETERMINING CALIBRATION FACTOR

$Cf^1$  . . . determined by using the 10 MHz out of the counter.

NOTE: At least 10 periods of the calibrating signal should fall within the external gate being used.  
In this case the period of 10 MHz = 100 nsec so 10 periods will be present in 1  $\mu$  sec.

$$Cf^1 = \frac{\text{CW value (INT GATE)}}{\text{pulsed value (EXT GATE)}} = \frac{10.0000}{9.9973} = 1.000270$$

$$\begin{aligned} \text{corrected video} &= F_{\text{VIDEO}} \times Cf = 345.57 \times 1.000270 \\ &= 345.66 \text{ MHz} \end{aligned}$$

Therefore, in this example we have an accuracy improvement factor over the single shot case of

$$\frac{\frac{0.2\%}{1 \times 100\%}}{\frac{0.2}{34566}} = \frac{0.2}{0.00289} = 69$$

As mentioned before, calibrating with the 10 MHz STD OUT signal, as above, will give an expected accuracy improvement factor of at least X20 over the single shot accuracy.

#### B. EVENTS MODE OF OPERATION

For Events Mode of Operation, system resolution and accuracy can be computed as follows:

$$1. \text{ Resolution} = \pm \left[ \frac{2 \text{ nsec}}{T} \times \text{freq} \right] \pm \left[ \frac{\text{trigger error}}{T} \times \text{freq} \right]$$

Where:

T = single period of event. Program is set for an external gate to the 5345A equal to the period of a particular event. If a larger external gate is desired, change line 11 of program from W1 to whatever width is desired.

$$\text{trigger error (rms)} = 1.4 \times \frac{\sqrt{X^2 + en^2}}{\text{Input slew rate at trigger point}}$$

X = 300  $\mu$  V for 5345A. Effective rms noise of counter's input channel.

en = rms noise voltage of input signal measured over a bandwidth equal to the counter's bandwidth.

$$2. \text{ Accuracy (Hz)} = \pm \text{Resolution} \pm [5345A \text{ time base error} \times \text{freq}]^* \pm \text{Bias}^*$$

\*5345A time base error and Bias are explained in the Time Mode of Operation.

## PROGRAM LISTINGS

A system measurement flow chart is shown in Figure 7. Typical system output listing is shown in Figure 8 and a program listing is shown in Figure 9.




Figure 7  
System Measurement Flowchart

FREQUENCY  
 PROFILE USING P  
 53458 COUNTER &  
 53598 TIME SYN-  
 THEORIZER

-----  
 TIME OPERATION  
 -----

SUCCESSIVE  
 WINDOW FREQUENCY  
 PROFILE 1  
 SET 53458:

Level H-Preset  
 Input 2-1Mohm  
 Attenu-X10  
 Coupling-DC  
 CHK/COMB/Sep-Sep

53598:  
 Width(sec)  
 1.000E-03  
 Delay(sec)  
 0.000E 00  
 Step size(sec)  
 1.000E-03

MANUAL SWEEP  
 Step Up 1  
 Freq 9.76732E 03  
 Step Up 2  
 Freq 9.76730E 03  
 Step Up 3  
 Freq 9.76741E 03  
 Step Down 1  
 Freq 9.76720E 03  
 Step Down 2  
 Freq 9.76725E 03  
 Step Down 3  
 Freq 9.76754E 03  
 AUTO SWEEP UP  
 Step Up 1  
 Freq 9.76708E 03  
 Step Up 2  
 Freq 9.76686E 03

Step up  
 Freq 9.76680E 03  
 AUTO SWEEP DOWN  
 Step Down 1  
 Freq 9.76603E 03  
 Step Down 2  
 Freq 9.76654E 03  
 Step Down 3  
 Freq 9.76701E 03  
 -----

EVENT OPERATION  
 -----

SUCCESSIVE  
 PERIOD FREQUENCY  
 PROFILE

53598:  
 Connect input  
 signal to read  
 panel EVENTS BNC

EventNumber 2  
 (+slope)  
 Per 1.02384E-04  
 Freq 9.77190E 03

EventNumber 5  
 Per 1.02374E-04  
 Freq 9.76290E 03

EventNumber 7  
 Per 1.02056E-04  
 Freq 9.76920E 03

EventNumber 10  
 Per 1.02384E-04  
 Freq 9.76980E 03

END

Figure 8

Figure 9

```

0: dev "5345",710;dev "5359",704
1: time 30000
2: dim AS[10],BS[10];flt 3;sfg 14
3: fmt 1,"W",e10.3
4: fmt 2,"D",e10.3
5: fmt 3,"SS",e10.3
6: fmt 4,"Step Up",f2.0
7: fmt 5,"Freq",e12.5
8: fmt 6,"Step Down",f2.0
9: fmt 7,"Per ",e12.5
10: fmt 8,"EventNumber",f5.0
11: fmt 9,"D",f2.0,"W1",",",f2
12: prt "      FREQUENCY","PROFILES USING A","5345A COUNTER &"
13: prt "5359A TIME SYN-","THESIZER";sps 2
14: ent "TIME(t)/EVENT(e) OPERATION?",A$;if A$="e";qto 74
15: spc 1;prt "-----";prt " TIME OPERATION";prt "-----"
16: prt "      SUCCESSIVE","WINDOW FREQUENCY","      PROFILE"
17: prt "SET 5345A:"," Level A-Preset"," Input Z-1Mohm"
18: prt "      Atten-X10","      Coupling-DC","Chk/ComA/Sep-Sep";sps 1
19: dsp "SELECT 5359 SETTINGS--Press CONT";sto
20: prt "5359A:"
21: ent "WIDTH XXX.XXe+/-X",A
22: prt "Width(sec)",A
23: if A>1.6e-1;beep;qto 21
24: if A<5e-9;beep;qto 21
25: wrt "5359.1",A
26: ent "DELAY XXX.XXe+/-X",B
27: prt "Delay(sec)",B
28: if B>1.6e-1;beep;qto 26
29: if A+B>1.6e-1;beep;qto 26
30: wrt "5359.2",B
31: ent "STEP SIZE XXX.XXe+/-X",C
32: prt "Step Size(sec)",C;sps 1
33: if C<5e-11;beep;qto 31
34: wrt "5359.3",C
35: wrt "5345","I2E";98"
36: ent "MAN(m)/AUTO(a)SWEEP? PressCONT",AS
37: if A$="a";qto 54
38: prt "MANUAL SWEEP";dsp "MANUAL SWEEP";wait 1000
39: l+N
40: ent "DELAY STEP UP?(yes=y,no=n)",A$
41: if A$="n";qto 46;l+N
42: wrt "5359","dsu";wait 200;wrt "5345","J1";wait 200
43: on err "CNTR"
44: red "5345",A;wrt 16.4,N;wrt 16.5,A
45: N+1+N;qto 40
46: l+M
47: ent "DELAY STEP DOWN?(yes=y,no=n)",A$
48: if A$="n";qto 53
49: wrt "5359","dsd";wait 200;wrt "5345","J1";wait 200
50: on err "CNTR"
51: red "5345",A;wrt 16.6,M;wrt 16.5,A
52: M+1+M;qto 47
53: ent "AUTO SWEEP?(yes=y,no=n)",A$;if A$="n";qto 73
54: prt "AUTO SWEEP UP";dsp "AUTO SWEEP UP";wait 1000
55: ent "No of Steps?(X) PressCONT",J
56: l+N
57: for I=1 to J
58: wrt "5359","dsu";wait 200;wrt "5345","J1";wait 200
59: on err "CNTR"
60: red "5345",A;wrt 16.4,N;wrt 16.5,A
61: N+1+N
62: next I
63: ent "AUTO SWEEP DOWN?(yes=y,no=n) ",A$;if A$="n";qto 73
64: prt "AUTO SWEEP DOWN"
65: ent "No of Steps?(X) PressCONT",J
66: l+N

```

```

67: for I=1 to J
68: wrt "5359", "dsp", wait 200; wrt "5345", "J1", wait 200
69: on err "CNTR"
70: red "5345", A; wrt 16.6, N; wrt 16.5, A
71: N+1+N
72: next I
73: ent "MANUAL SWEEP? (yes=y, no=n)", A$; if A$="y"; goto 38
74: spc 1; prt "-----"; prt "EVENT OPERATION"; prt "-----"
75: spc 1; prt "SUCCESSIVE", "PERIOD FREQUENCY", "PROFILE"; spc 1
76: prt "5359A:", "Connect input", "signal to rear", "panel EVENTS BNC"; spc 2
77: dsp "WHEN READY PRESS CONTINUE"; stp
78: wrt "5345", "I2G5FLE; E1E<E8"; wait 500
79: wrt "5359", "D2,W1,EP"
80: 2+N
81: red "5345", X; red "5345", A
82: wrt "5345", "F0"; wait 500
83: red "5345", X; red "5345", B
84: wrt 16.8, N; prt " (+100e)"; wrt 16.7, A; wrt 16.5, B; spc 2
85: ent "EVENT NUMBER?", N
86: if N+1>16777215; tmo -1
87: ent "EVENT SLOPE? (+,-)", BS
88: if BS="+", "EP"+BS
89: if BS="-", "EN"+BS
90: wrt "5345", "F1"; wait 500
91: wrt "5359.9", N, BS
92: red "5345", X; red "5345", A
93: wrt "5345", "F0"; wait 1000
94: red "5345", X; red "5345", B
95: wrt 16.8, N; wrt 16.7, A; wrt 16.5, B; spc 2
96: ent "ANOTHER MEASUREMENT? (yes=y, no=n)", A$; if A$="y"; goto 85
97: prt "END"; dsp "END"
98: end
99: "CNTR": beeo; dsp "NO 5345 MEASUREMENT"; wait 2000; goto 36
*2504

```